



## URGENT FIELD SAFETY NOTICE

<b>ATTENTION</b>	Central Sterilisation Department
<b>TYPE OF ACTION</b>	Voluntary FIELD SAFETY NOTICE
<b>REF / DATE</b>	ASP08/2013/October 02, 2013
<b>PRODUCT</b>	STERRAD <sup>®</sup> Boosters and Adaptors, Product Codes 15400, 15401, 15402, 15403 and 15407
<b>REASON</b>	The purpose of this communication is to announce the introduction of a new verification tool that should be used with your STERRAD <sup>®</sup> Boosters and STERRAD <sup>®</sup> Adaptors. This new tool is a checklist (Attached) that should be printed and used to provide verification of the proper use of the STERRAD <sup>®</sup> Booster/Adaptor before and after the STERRAD <sup>®</sup> sterilization cycle. (Note: This checklist should not be placed inside of the STERRAD <sup>®</sup> System or in any type of sterilization packaging)
<b>ACTION</b>	<b>What Action Is Required?</b> <ul style="list-style-type: none"><li>This new tool is a checklist (Attached) that should be printed and used to provide verification of the proper use of the STERRAD<sup>®</sup> Booster/Adaptor before and after the STERRAD<sup>®</sup> sterilization cycle.</li></ul>
<b>TRANSMISSION</b>	<b>Keep a copy of this letter with your STERRAD<sup>®</sup> Booster Instructions for Use (IFU) until you receive the updated IFU (Rev 104421-02) with your next STERRAD<sup>®</sup> Booster order.</b>
<b>CONTACT</b>	If you have additional questions about this action, please contact your Sales Representative or call [Affiliate Name].  We apologize for any inconvenience this will cause you, but rest assured it is our utmost intent to make this process as easy for you as possible.
<b>CONFIRMATION</b>	This action has been notified to the appropriate Regulatory Agencies.