



«Nom\_de\_Etablissement\_»  
«Nom\_du\_correspondant»

«Adresse\_1»  
«Adresse\_2»  
«CP\_VILLE\_»

Quality contact : ☎ 00 33 139 926 351  
✉ [nusur@vygon.com](mailto:nusur@vygon.com)

Mail ref : 14 1556 «N\_de\_Chrono»CO/NU Ecouen, November 12, 2014  
**Objet : Safety information - Lot recall**  
**Vyctic 3-way stopcock code 876.20 and 70876.20**

Dear Madam, Sir,

During a quality control check, a non-systematic punching defect of the paper side of the primary packaging was identified for certain lots of **Vyctic 3-way stopcock code 876.20 and 70876.20**.

The following lots of **VYCLIC code 876.20 or code 70876.20 are affected:**

CODES :	876.20	70876.20
LOTS :	030914FC	110914FC
	080914FC	150914FC
	230914FC	180914FC
	260914FC	220914FC
	300914FC	

**Please do not use these products and we ask you to withdraw them.**

We ask you to kindly acknowledge receipt of this letter, to complete and return the attached form by indicating for each code / lot involved the quantities withdrawn from your institution.

The French health authorities (ANSM) are informed of this recall.

For further information, you can directly contact M. Martin, quality contact for France, 06.11.85.17.92 or M. Hercouet, sales administration coordinator for France 01 09 92 63 35.

We apologize for the inconvenience and please accept the assurances of our highest consideration.

**Christine OBER - Pharmacist**  
**Postmarket Quality Director**

Service Qualité  
Tél : 01.39.92.63.51  
Fax : 01.39.92.64.82  
E-mail : [quality@vygon.com](mailto:quality@vygon.com)

Service Clients France  
Tél : 01.39.92.63.81  
Fax : 01.39.90.99.37  
E-mail : [commandes@vygon.com](mailto:commandes@vygon.com)

Service Clients Export  
Tél : 01.39.92.64.17  
Fax : 01.34.29.19.34  
E-mail : [export@vygon.com](mailto:export@vygon.com)





RECALL VYCLIC CODE 876.20 et 70876.20

**ACKNOWLEDGMENT AND CUSTOMER RESPONSE FORM**

Please complete and return this form:  
by fax: 01 39 92 64 82  
or e-mail to : nusur@vygon.com

**Name and address of the institution :**

**Full name of the person to contact:** 14«N\_de\_Chrono»CO/NU «Email\_correspondant»  
  
«CP\_VILLE\_»

**Function :**

**Phone number :**

**Please tick the appropriate boxes :**

**You have the Vyclic 3-way stopcock code 876.20:**

- ✓ Lot 030914FC     Yes    NO    If you have units, number of units removed: \_\_\_\_\_
- ✓ Lot 080914FC     YES    NO    If you have units, number of units removed: \_\_\_\_\_
- ✓ Lot 230914FC     YES    NO    If you have units, number of units removed: \_\_\_\_\_
- ✓ Lot 260914FC     YES    NO    If you have units, number of units removed: \_\_\_\_\_
- ✓ Lot 300914FC     YES    NO    If you have units, number of units removed: \_\_\_\_\_

**You have the Vyclic 3-way stopcock code 70876.20 :**

- ✓ Lot 110914FC     YES    NO    If you have units, number of units removed: \_\_\_\_\_
- ✓ Lot 150914FC     YES    NO    If you have units, number of units removed: \_\_\_\_\_
- ✓ Lot 180914FC     YES    NO    If you have units, number of units removed: \_\_\_\_\_
- ✓ Lot 220914FC     YES    NO    If you have units, number of units removed: \_\_\_\_\_

Signature and Date :

Service Qualité  
Tél : 01.39.92.63.51  
Fax : 01.39.92.64.82  
E-mail : quality@vygon.com

Service Clients France  
Tél : 01.39.92.63.81  
Fax : 01.39.90.29.37  
E-mail : commandes@vygon.com

Service Clients Export  
Tél : 01.39.92.64.17  
Fax : 01.34.29.19.34  
E-mail : export@vygon.com

