Luzernstrasse 21 4528 Zuchwil Switzerland Tel. +41 32 720 40 60 Fax +41 32 720 40 61 http://www.depuysynthes.com



To the ATTENTION of: Operating Room Manager

6 March 2015

URGENT NOTICE: MEDICAL DEVICE FIELD SAFETY NOTICE – FSN2014525 Pedicle Awls: Technique Guides & Outer Labels

| Technique Guide Description | Technique Guide Number | | |
|---|------------------------|--|--|
| Pangea Spine System, Technique Guide | 036.000.080 | | |
| USS II Polyaxial Spine System, Technique Guide | 036.000.216 | | |
| MATRIX Spine System – Deformity, Technique Guide | 036.001.180 | | |
| MATRIX Spine System – Degenerative, Technique Guide | 036.001.185 | | |
| Affected Labeling Description | Product Code | | |
| Pedicle Awl 4.0mm w/Canevasit Handle | 388.550 | | |
| Pedicle Awl 4.0mm w/Silicone Handle | 388.656 | | |

Dear Sir/Madam,

Synthes GmbH is initiating a voluntary medical device Field Safety Notice for the Surgical Technique Guides for Pangea, USS II Polyaxial, Matrix Deformity, and Matrix Degenerative systems as well as outer labels for two Pedicle Awls (388.550 and 388.656).

Our records indicate that you may have inventory that is impacted by this Field Safety Notice.

Reason for this Field Safety Notice:

Technique Guides for these four systems describe the Pedicle Awls as 4.0mm for use with screws (4.0mm – 7.0mm). These awls should be used with 5.0mm – 7.0mm screws. This Field Safety Notice is being issued in order to communicate changes being made to Technique Guides and outer labels in order to clarify the intended use of these two awls within these systems. There has been no change to the design or manufacture of the Pedicle Awls.

The revised Surgical Technique Guides contain the following corrections related to this Field Safety Notice:

| | Pangea Spine System: Technique Guide 036.000.080, Pedicle Awl 388.656 | | | |
|------|---|---|--|--|
| Page | Current | Revision | | |
| 8 | Pedicle Awl 4.0mm with Silicone | Pedicle Awl 4.0mm with Silicone Handle, length | | |
| | Handle, for Pedicle Screws | 255mm, for Pedicle Screws 5.0 to 7.0mm | | |
| 58 | Pedicle Awl 4.0mm with Silicone Handle, length 255mm, for Pedicle Screws 4.0 to 7.0mm | Pedicle Awl 4.0mm with Silicone Handle, length 255mm, for Pedicle Screws 5.0 to 7.0mm | | |



| USS II Polyaxial Spine System: Technique Guide 036.000.216, Pedicle Awl 388.550 | | | |
|---|---------------------------|--|--|
| Page | Current | Revision | |
| 10 | Pedicle Awl, length 230mm | Pedicle Awl 4.mm with Canevasit Handle, length 230mm_for Pedicle Screws 5.0 to 7.0mm | |

| MATRIX Spine System – Deformity: Technique Guide 036.001.180, Pedicle Awl 388.656 | | | |
|---|---|---|--|
| Page | Current | Revision | |
| 10 | Pedicle Awl 4.0mm with Silicone Handle, length 255mm, for Pedicle Screws 4.0 to 7.0mm | Pedicle Awl 4.0mm with Silicone Handle, length 255mm, for Pedicle Screws 5.0 to 7.0mm | |

| MATRIX Spine System – Degenerative: Technique Guide 036.001.185, Pedicle Awl 388.656 | | | |
|--|---|---|--|
| Page | Current | Revision | |
| 8 | Pedicle Awl 4.0mm with Silicone Handle, length 255mm, for Pedicle Screws 4.0 to 7.0mm | Pedicle Awl 4.0mm with Silicone Handle, length 255mm, for Pedicle Screws 5.0 to 7.0mm | |
| 66 | Pedicle Awl 4.0mm with Silicone Handle, length 255mm, for Pedicle Screws 4.0 to 7.0mm | Pedicle Awl 4.0mm with Silicone Handle, length 255mm, for Pedicle Screws 5.0 to 7.0mm | |

^{*}All applicable translations will also be updated and made available accordingly.

The revised outer labels contain the following corrections related to this Field Safety Notice:

| Product Label | Current | Revision | | |
|------------------|--|----------|--|--|
| 388.550 | Pedicle Awl 4.0mm w/Canevasit Handle, L 230mm, f/Pedic. Screws 4.0 to 7.0mm | | | |
| 388.656 | Pedicle Awl 4.0mm w/Silicone Handle, L 255mm, f/Pedic.Screws 4.0 to 7.0mm | | | |

Please note that there has been no change to the design or manufacturing of these Pedicle Awls. This Field Safety Notice impacts information in Technique Guides and outer labels only.

Clinical Implications:

In the event that a Ø4mm Awl is used for a Ø4mm screw, there is a risk that perforation of the cortex will be too wide and may result in screw pull-out or loosening. In the event that the cortex is perforated wider than needed for insertion of a 4mm screw, the failure would not typically be recognized. Cases may occur, however, where a surgeon's perception of torque to insert the screw does not have the desired level of fixation. In this case, the issue may be recognized.

Please be advised that Technique Guides containing the updates referenced above will be available at www.Synthes.com as is customary for Synthes GmbH systems. Your facility may temporarily receive Pedicle Awls without the above referenced labeling clarifications due to their previous release to distribution channels. As we work to implement these labeling updates in our inventory, please note that there have been no updates made to the design or manufacture of these instruments. This Field Safety Notice impacts information in Technique Guides and outer labels only.

To date, there have been zero reports of related failure modes with the 4.0mm screws in these four systems that would lead to the aforementioned harms.



Customer immediate actions:

- 1. Review the revisions being made to surgical Technique Guides and outer labels.
- 2. Forward this notice to anyone in your facility that needs to be informed.
- 3. If any of the affected Technique Guides or outer labels have been forwarded to another facility, contact that facility to inform them of revisions.
- 4. Maintain awareness of this notice until all Technique Guides and outer labels have been updated.
- 5. Keep a copy of this notice.

Applicable Regulatory Agencies will be notified.

We apologize for any inconvenience that this Field Safety Notice may create and appreciate your attention to this matter. Should you have any inquiries, please do not hesitate to contact your DePuy Synthes sales consultant.

Thank you for your attention and cooperation.

Pierre VAN IWAARDEN.

Synthes GmbH

Pierre van Iwaarden

Field Action Manager

Charles Goldberg

WW Director, Complaint Management

Cc:



URGENT NOTICE: MEDICAL DEVICE FIELD SAFETY NOTICE – FSN2014525 Pedicle Awls: Technique Guides & Outer Labels

| Technique Guide Description | Technique Guide Number | | |
|---|------------------------|--|--|
| Pangea Spine System, Technique Guide | 036.000.080 | | |
| USS II Polyaxial Spine System, Technique Guide | 036.000.216 | | |
| MATRIX Spine System – Deformity, Technique Guide | 036.001.180 | | |
| MATRIX Spine System – Degenerative, Technique Guide | 036.001.185 | | |
| Affected Labeling Description | Product Code | | |
| Pedicle Awl 4.0mm w/Canevasit Handle | 388.550 | | |
| Pedicle Awl 4.0mm w/Silicone Handle | 388.656 | | |

Please note that there has been no change to the design or manufacturing of these Pedicle Awls. This Field Safety Notice impacts information in Technique Guides and outer labels only.

| We acknowledge the receipt of this i 4.0mm w/Canevasit Handle and/or this facility. | | | • | |
|---|-----------------|--------------|---------------------|-------------|
| We acknowledge the receipt of this i 4.0mm w/Canevasit Handle and/or this facility. | | | er Francisco esta e | |
| Name/Title (please print): | | 3 0 | | |
| Address: | | | | |
| Phone Number: | | | | |
| Signature and Date: | , | | | e e = 7 |
| Please complete and return this page y | our local DePuy | Synthes sale | s organization | 1. |

Note: If the Verification Section is answered on behalf of more than one facility and/or individual, please clearly indicate the name and address of the facility and/or individual on this page of the notification.