



URGENT MEDICAL DEVICE RECALL NOTICE
November 17, 2015

Integral Lateralized Femoral Stems
Part Numbers: reference chart on page 2
Lots: reference chart on page 2

Dear Risk/Recall Manager,

This notification is to inform you of an Urgent Medical Device Recall initiated by Zimmer Biomet Orthopedics ("Zimmer Biomet") which involves specific **Integral Lateralized Femoral Stems** (reference page 2 for specific part numbers and lot numbers).

These implant(s) have been consigned and/or invoiced to your account. Zimmer Biomet has initiated this action following an investigation which identified that PMMA is listed as a material on the label but the product does not contain PMMA.

The specific part numbers and lot numbers that have been consigned and/or invoiced to your account are listed on page 2.

This action requires the immediate location and discontinued use of the product and its return to Zimmer Biomet.

Adverse reactions or quality problems experienced with the use of this product may be reported to the FDA:

- Online: www.fda.gov/Safety/MedWatch/HowToReport/default.htm (form available to fax or mail), or
- Call (800)FDA-1088

Thank you in advance for your assistance and prompt attention. On behalf of Zimmer Biomet, I apologize for any inconvenience this may cause. Questions related to this notice should be directed to (574) 372-1570, Monday through Friday, 8 a.m. to 5 p.m.

Sincerely,

A handwritten signature in black ink that reads 'Audrey E. Daenzer'.

Audrey Daenzer
Field Action Specialist
audrey.daenzer@zimmerbiomet.com

Mailing Address:	Shipping Address:
P.O. Box 587	56 E Bell Drive
Warsaw, IN. 46581-0587	Warsaw, IN 46582
Toll Free: 800.348.9500	
Office: 574.267.6639	
Main Fax: 574-267-8137	
www.zimmerbiomet.com	



URGENT MEDICAL DEVICE RECALL NOTICE
November 17, 2015

All part numbers and lots numbers affected by this recall are listed below.

Part Number	Lot Number	Description
11-162709	329750	INTGL COCR LAT CENTRALR 9MM
11-162709	404380	INTGL COCR LAT CENTRALR 9MM
11-162709	487400	INTGL COCR LAT CENTRALR 9MM
11-162709	767120	INTGL COCR LAT CENTRALR 9MM
11-162709	951680	INTGL COCR LAT CENTRALR 9MM
11-162711	329760	INTGL COCR LAT CENTRALR 11MM
11-162711	505030	INTGL COCR LAT CENTRALR 11MM
11-162711	505050	INTGL COCR LAT CENTRALR 11MM
11-162713	837750	INTGL COCR LAT CENTRALR 13MM
11-162713	977460	INTGL COCR LAT CENTRALR 13MM
11-162715	772410	INTGL COCR LAT CENTRALR 15MM
11-162715	837780	INTGL COCR LAT CENTRALR 15MM
11-162715	897070	INTGL COCR LAT CENTRALR 15MM
11-162715	951810	INTGL COCR LAT CENTRALR 15MM
11-162717	311410	INTGL COCR LAT CENTRALR 17MM
11-162717	410500	INTGL COCR LAT CENTRALR 17MM
11-162717	535530	INTGL COCR LAT CENTRALR 17MM
11-162717	712800	INTGL COCR LAT CENTRALR 17MM
11-162717	728080	INTGL COCR LAT CENTRALR 17MM
11-162717	829580	INTGL COCR LAT CENTRALR 17MM

Mailing Address:
P.O. Box 587
Warsaw, IN. 46581-0587
Toll Free: 800.348.9500
Office: 574.267.6639
Main Fax: 574-267-8137
www.zimmerbiomet.com

Shipping Address:
56 E Bell Drive
Warsaw, IN 46582