

**Urgent!**  
**Field Safety Notice (FSN)**

22. August 2016

<b>FSN Number:</b>	2016-07-22
<b>FSN Title:</b>	Revised decontamination procedure for Heater- and Heater-Cooler Units
<b>Type of Action:</b>	Customer Notification
<b>Affected Products:</b>	Heater Unit HU 35 Heater-Cooler Unit HCU 20 Heater-Cooler Unit HCU 30 Heater-Cooler Unit HCU 40
<b>Unique Device Identification (UDI Code):</b>	Not assigned
<b>Affected product details:</b>	The FSN affects all models and all serial numbers of the above mentioned devices
<b>Description of the problem:</b>	<p>Maquet Cardiopulmonary as well as the National Competent Authorities in multiple countries have received reports confirming the presence of Mycobacterial contamination (and other bacterial contamination) in the system water of Heater and Heater-Cooler devices, mainly used during cardiothoracic surgeries to warm or cool a patient.</p> <p>Although the water in the circuits does not come into direct contact with the patient, there is the potential for contaminated water to enter other parts of the device or transmit bacteria through the air (aerosolize) through the device's exhaust vent into the environment and to the patient.</p> <p>Maquet Cardiopulmonary has confirmed that Maquet systems used in these procedures might be contaminated with various strains of bacteria that also include Mycobacteria species.</p> <p>Although there is a risk that the contamination can be transferred to patients, Maquet Cardiopulmonary has not received any reports that Mycobacteria contamination has been confirmed in a patient treated with HCU 40, HCU 30, or HCU 20. We have also not received any reports for those units concerning adverse events, diseases, illness associated with Mycobacteria contaminations.</p>

**Corrective Action:**

In case of confirmed bacterial contaminations of Maquet HCUs, we recommend to take the devices affected out of operation at the earliest opportunity and emptying the system water with a subsequent disinfection according to the current protocol. The re-commissioning of the device should be supported by regular hygienic monitoring and as circumstances require a shorter disinfection interval.

Preventive measures of within the framework of common and recommended hospital hygiene reduce significantly the risks of a bacterial transmission after contact with contaminated system water. These measures include:

- the use of gloves
- sterile filter for filling the units
- proper hygienic monitoring
- regular water changes and
- disinfection of the units

Maquet Cardiopulmonary is currently developing new hygiene protocols for its Heater-Cooler units HCU 40, HCU 30 and HCU 20 as well as the Heater Unit HU 35. These new protocols include preventive measures, routine disinfection as well as high level disinfection and biofilm reduction – also effective against atypical Mycobacteria in the water systems. The newly developed disinfection procedures are taking a holistic approach based on validated methods.

Moreover, we implemented an effective hygienic monitoring process in our production. 100% microbiological control of every produced HCU 40 is guaranteed for every unit before leaving the factory.

The current time schedule for the introduction of the new hygiene protocols:

Heater Cooler Unit HCU40:	2016-11-30
Heater Cooler Unit HCU30:	2016-12-15
Heater Cooler Unit HCU20:	2016-12-15
Heater Unit HU35:	2016-11-15

**Advice on action to be taken by the user:**

Complete and return the enclosed Acknowledgement Form after reading to your local Maquet representative.

User should continue to follow the Instructions for Use received with the devices.

Continue to monitor the hygiene (contamination level) in accordance with your internal practices.

Immediately report any contamination finds to your local Maquet representative.

**Referenced documents/attachments:**

Acknowledgement Form

### Transmission of the Field Safety Notice:

- This notice needs to be passed on to all those who need to be aware within your organization or to any organization where the potentially affected devices have been transferred.
- Please transfer this notice to other organizations on which the action has an impact.
- Please maintain awareness on the notice and resulting actions for an appropriate period to ensure effectiveness of the corrective action.

We apologize for any inconvenience this may cause you and we will do our utmost to carry through this action as swiftly as possible.

As required, we have provided this notification to the necessary Regulatory Agencies.

Should you have questions or require additional information, please contact your local Maquet representative.

Sincerely,



**Managing Director** Mikael K Johansson



**Safety Officer**

Maquet Cardiopulmonary GmbH  
Kehler Str. 31  
76437 Rastatt  
GERMANY