

**ADVIA Centaur®
 ADVIA Centaur® XP
 ADVIA Centaur® XPT
 ADVIA Centaur® CP**

**ADVIA Centaur Systems Calibrator U:
 Analytical Sensitivity Limit with the Myoglobin Assay**

Our records indicate that your facility may have received the following product:

Table 1. ADVIA Centaur Affected Product(s)

Product	Test Code	Catalog Number (REF)	Siemens Material Number (SMN)	Kit Lots Ending In	Expiration Date	Manufacturing Date
Calibrator U (Myoglobin)	Cal U	03684480	10309996	63 64	2016-10-03 2017-01-17	2015-12-03 2016-03-17

Reason for Correction

Siemens Healthcare Diagnostics has confirmed that the ADVIA Centaur Systems Myoglobin assay is not meeting the analytical sensitivity claim of ≤ 3 ng/mL (ug/L) as specified in the ADVIA Centaur Myoglobin Instructions for Use (IFU), for all in-date reagent lots when evaluated with Calibrator U kit lots ending in 63 and 64. Values up to 12 ng/mL (ug/L) were observed.

Siemens' investigation has identified that the Calibrator U kit lots listed in Table 1 above have drifted from the internal standardization causing a positive shift in results.

Analytical sensitivity and alignment to the internal standardization will be restored with the release of Calibrator U kit lots ending in 65 (CU65) and higher.

Customers will observe a negative shift in Quality Control (QC) material, Master Curve Material (MCM) and patient results when transitioning to CU65. (Refer to the Additional Information section of this letter for further details). As a result, revised QC targets and ranges for Bio-Rad controls have been established and published on the Bio-Rad website at QCnet.com and a new lot of MCMs, Lot 05125 (expiry 2018-06-11), will be made available.

Risk to Health

Positively biased serum myoglobin values approximating the clinical decision point may lead to additional follow-up and/or investigations to assess acute muscle damage with negligible risk to health. Serum myoglobin is not used in isolation for diagnosing rhabdomyolysis or cardiac injury. Elevated values for serum myoglobin would be interpreted in the context of clinical history/presentation as well as with other biomarkers such as troponin and/or creatine kinase.

Siemens is not recommending a review of previously generated results.

Actions to be Taken by the Customer

- Please review this letter with your Medical Director.
- Customers may continue to use Calibrator U lots CU63 and CU64 with any in-date reagent kit lot, current QC ranges and current Master Curve Material lot 45719.
- Revised QC targets and ranges are located on the Bio-Rad website at QCnet.com for use with CU65 and higher.
- Myoglobin Master Curve Material Lot 05125 (expiry 2018-06-11) is to be used with CU65 or higher and should not be used with Calibrator U lot CU63 or CU64.
- Review the information provided in the Additional Information section of this communication.
- Complete and return the Field Correction Effectiveness Check attached to this letter within 30 days.

Please retain this letter with your laboratory records, and forward this letter to those who may have received this product.

We apologize for the inconvenience this situation may cause. If you have any questions, please contact your Siemens Customer Care Center or your local Siemens technical support representative.

Product availability may vary from country to country and is subject to varying regulatory requirements. Due to local regulations, the ADVIA Centaur XPT is not available in all countries.

Additional Information

Based on Siemens' internal testing, the following biases may be observed in QC and patient results when transitioning to CU65. Biases, by dose range, between Calibrator Lots CU65 and CU64 are listed in Table 2 below.

Table 2. Bias Comparison CU65 to CU64

Dose Range (ng/mL) (ug/L)	% Bias					
	ADVIA Centaur/XP/XPT Systems			ADVIA Centaur CP System		
	Mean	Max Bias	Min Bias	Mean	Max Bias	Min Bias
<20	-17.1%	-18.9%	-15.4%	-43.5	-48.3%	-38.7%
21-60	-10.7%	-14.3%	-7.92%	-23.5%	-34.8%	-13.9%
61-100	-7.33%	-7.74%	-6.53%	-11.9%	-12.7%	-9.22%
101-110	-6.16%	N/A*	N/A*	-7.46%	-7.56%	-7.36%
111-150	-5.78%	-6.04%	-5.61%	-5.75%	-6.44%	-5.34%
151-200	-5.43%	-5.60%	-5.25%	-4.61%	-5.21%	-3.92%
201-1000	-4.70%	-5.25%	-4.43%	-1.49%	-3.87%	-0.44%

* Not calculated since only one result within this dose range.

Similar results are expected to be observed between Calibrator Lots CU65 and CU63.

Figure 1 and Figure 2 provide linear regression comparing CU65 to CU64. The data was generated with ADVIA Centaur Myoglobin reagent kit lot ending in 187.

Figure 1: ADVIA Centaur/XP/XPT Patient Correlation: CU65 vs. CU64





Figure 2: ADVIA Centaur CP Patient Correlation: CU65 vs. CU64



ADVIA Centaur is a trademark of Siemens Healthcare Diagnostics.

FIELD CORRECTION EFFECTIVENESS CHECK

ADVIA Centaur Systems Calibrator U:
Analytical Sensitivity Limit with the Myoglobin Assay

This response form is to confirm receipt of the enclosed Siemens Healthcare Diagnostics Urgent Field Safety Notice CC 16-18.A.OUS dated August, 2016 regarding ADVIA Centaur Systems Calibrator U: Analytical Sensitivity Limit with the Myoglobin Assay. Please read each question and indicate the appropriate answer. Fax this completed form to Siemens Healthcare Diagnostics at the fax number provided at the bottom of this page.

1. I have read and understood the Urgent Field Safety Notice instructions provided in this letter. Yes No

Name of person completing questionnaire: _____

Title: _____

Institution: _____ Instrument Serial Number: _____

Street: _____

City: _____ State: _____

Phone: _____ Country: _____

Please fax this completed form to the Customer Care Center at (XXX) XXX-XXXX. If you have any questions, contact your local Siemens technical support representative.