



FSN Ref: 2019-12 (01)
Date: DD.MMM.2019.

FSCA Ref: 2019-12 (01)

Urgent Field Safety Notice
Mölnlycke® Procedure Trays

For Attention of: Theatre Manager

Contact details of local representative (name, e-mail, telephone, address etc.)
Name: Local Customer Care contact will be added for each specific market Email: XXX.XXX@mölnlycke.com Telephone: +XXXXXXXXXXXXXXXXX

Urgent Field Safety Notice (FSN)
Mölnlycke® Procedure Trays
Latex label error on the Insert Card

1. Information on Affected Devices	
1.	<p>1. Device Type(s)</p> <p>Mölnlycke® Procedure Trays consist of customized configurations of components, which are assembled and delivered sterile within one packaging.</p>
1.	<p>2. Commercial name(s)</p> <p>See Appendix I Product Table</p>
1.	<p>3. Primary clinical purpose of device(s)</p> <p>The clinical purpose of Mölnlycke® Procedure Trays is to provide a customized sterile co-packing of components for different clinical interventions.</p>
1.	<p>4. Device Model/Catalogue/part number(s)</p> <p>See Appendix I Product Table</p>
1.	<p>5. Affected serial or lot number range</p> <p>See Appendix I Product Table</p>

2 Reason for Field Safety Corrective Action (FSCA)	
2	<p>1. Description of the product problem*</p> <p>Mölnlycke® has identified a potential safety issue. During a regular inspection, an error with the label (Insert Card) was identified. The information on the label incorrectly states that the Mölnlycke® Procedure tray does not contain “Natural Rubber Latex”; however, some components within these trays do contain latex. The label should state that the tray does contain latex.</p>
2	<p>2. Hazard giving rise to the FSCA*</p> <p>If the concerned product is used by a latex sensitive person an allergic reaction may follow. Risk is associated to both user and patient. Please ensure that the affected Mölnlycke® Procedure trays are not used by/on anyone sensitive to latex.</p>

3. Type of Action to mitigate the risk	
3.	<p>1. Action To Be Taken by the User</p> <p><input checked="" type="checkbox"/> Identify Device</p> <p><input checked="" type="checkbox"/> Quarantine Device</p> <p><input checked="" type="checkbox"/> Take note of amendment: affix a copy of this FSN to each Mölnlycke® Procedure tray and make this FSN information available for the user</p> <p>We need your help in ensuring that <u>all affected products</u> are located and that below actions are performed.</p> <p>Please follow below instructions:</p>



FSN Ref: 2019-12 (01)

FSCA Ref: 2019-12 (01)

Date: DD.MMM.2019.

	<ol style="list-style-type: none">1. Identify and isolate the product at your facility, please see Appendix I for affected product information.2. Affix a copy of this Field Safety Notice (FSN) to each Mölnlycke® Procedure tray and make sure that its contents is brought to the attention of all relevant personnel to read before use.3. Fill out the Customer Reply Form, Appendix II, and return it back to Mölnlycke within 10 business days, even if you do not have affected products. Mölnlycke needs to be sure all customers are aware of the situation.4. If you have forwarded any affected products to other healthcare institutions, please send them a copy of this Field Safety Notice. Make sure they act accordingly.5. If you are a distributor, please inform your customers by sending them a copy of this Field Safety Notice. Make sure they act accordingly and return the Customer reply form in Appendix II to you. <p>We apologize for any inconvenience this will cause you, and rest assured it is our utmost intent to make this process as easy for you as possible.</p> <p>In addition, Mölnlycke appreciates your help in collecting data on product complaints and/or incidents related to the concerned product. Please follow the reporting procedures established by your facility.</p>	
3.	2. Is customer Reply Required?	Yes (Within 10 business days)

4. General Information		
4.	1. FSN Type	New
4.	2. Further advice or information already expected in follow-up FSN?	No
4.	3. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)	
	a. Company Name	Mölnlycke Health Care AB
	b. Address	Box 130 80, SE-402 52 Gothenburg, Sweden
	c. Website address	www.molnlycke.com
4.	4. The Competent (Regulatory) Authority of your country has been informed about this communication to customers.	
4.	5. List of attachments/appendices:	Appendix I Product table Appendix II Customer Reply Form
4.	6. Name/Signature	Linda Magnusson, Post Market Surveillance and Site Quality Director

Transmission of this Field Safety Notice	
	<p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.</p>

FSN Ref: 2019-12 (01)
Date: DD.MMM.2019.
Appendix I

FSCA Ref: 2019-12 (01)

Product table

To be added for each market

FSN Ref: 2019-12 (01)

FSCA Ref: 2019-12 (01)

Date: DD.MMM.2019.

Appendix II

Customer Reply Form

1. Field Safety Notice (FSN) information	
FSN Reference number	2019-12 (01)
FSN Date	DD.MMM.2019.
Product/ Device name	See Appendix I Product table
Product Code(s)	See Appendix I Product table
Batch/Serial Number (s)	See Appendix I Product table

2. Customer Details	
Account Number	
Healthcare Organisation Name*	
Organisation Address*	
Department/Unit	
Shipping address if different to above	
Contact Name*	
Title or Function	
Telephone number*	
Email*	

3. Customer action undertaken on behalf of Healthcare Organisation	
<input type="checkbox"/>	I confirm receipt of the Field Safety Notice and that I read and understood its content. The information and required actions have been brought to the attention of all relevant users and executed.
<input type="checkbox"/>	I confirm receipt of the Field Safety Notice and that I read and understood its content. I do not have any affected devices.
Print Name*	
Signature*	
Date*	

4. Return acknowledgement to sender	
Email	vigilance@molnlycke.com
Customer Helpline	+XXXXXXXXXXXXXXXX
Postal Address	Mölnlycke Health Care,

FSN Ref: 2019-12 (01)

FSCA Ref: 2019-12 (01)

Date: DD.MMM.2019.

	Box 130 80, SE-402 52 Gothenburg, Sweden
Fax	+46 31 722 34 00
Deadline for returning the customer reply form*	Within 10 days

Mandatory fields are marked with *

It is important that your organisation takes the actions detailed in the FSN and confirms that you have received the FSN.

Your organisation's reply is the evidence we need to monitor the progress of the corrective actions.