



Rev 1: September 2018
FSN Ref: 0020-1790

FSCA Ref: 0020-1791

Date: 06Mar2021

Urgent Field Safety Notice
Extreme H2O 59% Daily Contact Lenses

For Attention of*:Logistics Manager or Site Manager

Contact details of local representative (name, e-mail, telephone, address etc.)*

| Distributor | Address | City | Zip | Country | Phone |
|--------------------------------------|---------------------------|-----------------|-------|--------------------|-------------|
| Techlens WL Contactlinsen GmbH | Schleissheimer Str 267 | Munchen | 80809 | Germany | 49893236700 |
| Hydrolens - Vision Care | Nordlandsvej 86 | Risskov | 8240 | Denmark | 4586271766 |
| Galifa Contactlinsen | Zurcherstrasse 204e | Saint Gallen | 9014 | Switzerland | 41712723000 |
| Ercon Bv | Afrikaweg 51 | Assen | 9407 | The Netherlands | 31592405000 |

Please also feel free to contact Clerio Vision at +1-941-739-1382 between 9am and 5pm EST or email us at sarasotacustomer@cleriovision.com.

Urgent Field Safety Notice (FSN)
Extreme H2O 59% Daily Contact Lenses
Mislabeled Product

| 1. Information on Affected Devices* | | | | |
|-------------------------------------|---|---------------------|-------|------------|
| 1. | 1. Device Type(s)* | | | |
| | Sterile soft contact lens | | | |
| 1. | 2. Commercial name(s) | | | |
| | Extreme H2O 59% Daily Lens | | | |
| 1. | 3. Primary clinical purpose of device(s)* | | | |
| | Spherical soft contact lens for daily wear is indicated for the correction of visual acuity in aphakic or not-aphakic persons with non-diseased eyes that are myopic or hyperopic. The lens may be worn by persons who exhibit astigmatism of 0.75 Diopters or less that does not interfere with visual acuity. | | | |
| 1. | 4. Device Model/Catalogue/part number(s)* | | | |
| | UPC Code | Product | Power | Package |
| | 675506700657 | Extreme H2O 59 Xtra | +3.75 | 6 Pack |
| | 675506668650 | Extreme H2O 59 Xtra | +3.75 | Individual |
| 1. | 5. Affected serial or lot number range | | | |
| | Lot: 0114511565 | | | |

| 2 Reason for Field Safety Corrective Action (FSCA)* | |
|---|--|
| 2. | 1. Description of the product problem* |
| | It has been determined that Lot#:0114511565 had lenses that were accidentally mislabeled. The power of the lens printed on the label shows +3.75 Diopters when in fact, some of the lens in the blister packages are a -2.00 Diopter lens. |
| 2. | 2. Hazard giving rise to the FSCA* |
| | The only existing hazard is reduced visual acuity. Given that 5.75 diopter difference this issue would be immediately noticed by the patient and they would remove and not use the lens due to the lack of visual acuity. These lenses have been sterilized and are safe for use. They will only reduce visual acuity. |
| 2. | 3. Probability of problem arising |
| | 50% |
| 2. | 4. Predicted risk to patient/users |
| | No harm exists for the patient. Only a noticeable loss in visual acuity |
| 2. | 5. Background on Issue |
| | This issue was discovered by customer complaint. The root cause has been determined and all existing inventory segregated. Additional quality checks will be added to the process to prevent future occurrences. |

| 3. Type of Action to mitigate the risk* | |
|--|--|
| 3. | <p>1. Action To Be Taken by the User*</p> <p> <input type="checkbox"/> Identify Device <input type="checkbox"/> Quarantine Device <input checked="" type="checkbox"/> Return Device <input checked="" type="checkbox"/> Destroy Device <input type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> Follow patient management recommendations <input type="checkbox"/> Take note of amendment/reinforcement of Instructions For Use (IFU) <input type="checkbox"/> Other <input type="checkbox"/> None </p> <p>Provide further details of the action(s) identified.</p> |
| 3. | <p>2. By when should the action be completed?</p> <p>Not critical to safety but should be returned or destroyed with evidence of destruction returned to Clerio Vision as soon as possible.</p> |
| 3. | <p>3. Is customer Reply Required? * (If yes, form attached specifying deadline for return)</p> <p style="text-align: right;">Yes</p> |
| 3. | <p>4. Action Being Taken by the Manufacturer</p> <p> <input checked="" type="checkbox"/> Product Removal <input type="checkbox"/> On-site device modification/inspection <input type="checkbox"/> Software upgrade <input type="checkbox"/> IFU or labelling change <input type="checkbox"/> Other <input type="checkbox"/> None </p> <p>All affected product is being removed from the market and additional quality checks are being put in place to avoid future occurrence</p> |
| 3 | <p>5. By when should the action be completed?</p> <p>Not critical to safety but should be returned or destroyed with evidence of destruction returned to Clerio Vision as soon as possible.</p> |
| 3. | <p>6. Is the FSN required to be communicated to the patient /lay user?</p> <p style="text-align: right;">Yes</p> |
| 3 | <p>7. If yes, has manufacturer provided additional information suitable for the patient/lay user in a patient/lay or non-professional user information letter/sheet?</p> <p>Yes Appended to this FSN</p> |

| 4. General Information* | | |
|--------------------------------|--|---|
| 4. | 1. FSN Type* | New |
| 4. | 2. Further advice or information already expected in follow-up FSN? * | No |
| 4. | 3. Manufacturer information (For contact details of local representative refer to page 1 of this FSN) | |
| | a. Company Name | Clerio Vision |
| | b. Address | 7575 Commerce Ct. SARASOTA, FL 34243 USA |
| | c. Website address | https://extremeh2o.com/ |
| 4. | 4. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. Yes | |
| 4. | 5. List of attachments/appendices: | Customer Response |
| 4. | 6. Name/Signature | Ian Kelsey Quality Manager |
| | | |

| Transmission of this Field Safety Notice | |
|---|--|
| | <p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback..*</p> |

Note: Fields indicated by * are considered necessary for all FSNs. Others are optional.