



## Urgent Field Safety Notice

PH-21-003 A OUS

March 2021

### BCS® XP System

#### Potential carryover of Emicizumab by patient samples

---

Our records indicate that your facility may have received the following product:

**Table 1. BCS XP System Affected Product(s)**

System	Siemens Material Number
BCS XP System	10459330, 10461894, 10470625, 11240019

#### Reason for Correction

The purpose of this communication is to inform you about an issue with the product indicated in Table 1 above and provide instructions on actions that your laboratory must take.

Siemens Healthcare Diagnostics Products GmbH has confirmed that the BCS XP System may be affected by a potential Emicizumab carryover on patient samples.

Based on our investigations only the low Factor VIII clotting applications may be affected by this issue. Patient samples with Factor VIII concentrations above 15% of norm, measured in the normal Factor VIII clotting applications, are not affected.

Siemens Healthcare Diagnostics Products GmbH determined that an intensive washing after the measurement of an Emicizumab patient sample can eliminate the potential carryover. For further information, please see section "Additional Information".

#### Risk to Health

Given the fact that only samples measured with the low Factor VIII applications directly after a sample from a patient treated with Emicizumab are affected by the carryover, the likelihood of this combination has been assessed to be very low.

However, there is still a very low possibility of overestimating samples measured with the low Factor VIII applications of patients with Hemophilia A if a carryover occurred.

Due to the extremely low expected probability of occurrence no general lookback procedure is recommended. In individual cases, when a low F VIII deviates from an anticipated F VIII level (only with the low application), a confirmation test should be concluded.

## BCS XP System – Potential carryover of Emicizumab by patient samples

### Actions to be Taken by the Customer

- Please review this letter with your Medical Director.
- Perform the instructions provided in Additional Information.
- Please retain this letter with your laboratory records and forward this letter to those who may have received this product.

### Additional Information

We recommend to measure patient samples with known Emicizumab treatment in batch mode followed by the intensive washing cycle described below.

After the measurement of a potential Emicizumab patient sample an intensive washing cycle of the sample probe is recommended, prior measuring any further sample with a low Factor VIII clotting application.

The intensive washing cycle is using a washing solution (Washing Solution for Coagulation Analyzers) in addition to the regular water rinse for the sample probe.

This additional washing step must be performed manually via the user software as described in the BCS XP System Instruction Manual, Chapter 11.1.4. This washing cycle cannot be implemented into the assay procedure for an automatic use.

In addition, results of Hemophilia patients measured with any F VIII low application should be carefully reviewed considering clinical history. In case of doubt, remeasuring is recommended.

We apologize for the inconvenience this situation may cause. If you have any questions, please contact your Siemens Healthineers Customer Care Center or your local Siemens Healthineers technical support representative.

*This letter was created electronically and is valid without signature*

i. V. Dr. Norbert Dedner  
Senior Director  
Quality Systems & Compliance

i. A. Sandra Otto  
Product Manager  
Global Marketing

BCS is a trademark of Siemens Healthcare Diagnostics Products GmbH.