

Urgent Field Safety Notice

Endo Stitch™ Single Use Loading Unit with V-Loc™ Reload

Recall

September 2022

Medtronic Reference: FA1281

EU Manufacturer Single Registration Number (SRN): US-MF-000028763

Dear Risk Manager/Healthcare Professional,

The purpose of this letter is to advise you that Medtronic has initiated a voluntary recall for specific lots of the Endo Stitch™ Single Use Loading Unit with V-Loc™ 180 Absorbable Reload and the Endo Stitch™ Single Use Loading Unit with V-Loc™ PBT Non-Absorbable Reload. You are receiving this letter as Medtronic records indicate your facility may have at least one of the Endo Stitch Single Use Loading Unit with V-Loc Reloads identified for recall in Appendix 1. Medtronic initiated this action to prevent the use of potentially affected Endo Stitch Single Use Loading Unit with V-Loc Reloads that may impact patients.

Issue Description:

Medtronic has received reports of needles breaking during endoscopic suturing. The potential harm(s) include a delay in treatment/therapy, the potential for a portion of the needle to remain in the patient, foreign body reaction, allergic reaction, tissue injury and unintended radiation exposure from additional imaging for both intra-op and postoperative scenarios.

As of 13 September 2022, Medtronic has received 210 complaints related to this issue worldwide. Of these, 119 reported injuries potentially related to the needle breakage. These injuries include foreign body in patient, tissue damage, extended operating room time, and unintended radiation exposure. The cause of the issue is related to a manufacturing nonconformance, and process improvements are in progress to address the nonconformance prior to resuming manufacturing and distribution of product.

There are no additional patient management recommendations for patients where an Endo Stitch Single Use Loading Unit with V-Loc Reload in scope of this recall was used during a procedure. These patients should continue to be monitored in accordance with your medical facility's standard care protocols.

Medtronic

Product Scope:

Product Name	Models	Lot Numbers
Endo Stitch V-Loc™ 180 Absorbable Reload	VLOCA004L, VLOCA006L, VLOCA008L, VLOCA204L, VLOCA206L, VLOCA208L, VLOCA304L, VLOCA306L, VLOCA308L	Unexpired lots beginning with prefix "N0", "N1", "N9", "N2A", "N2B", "N2C", "N2D", and "N2E"
Endo Stitch V-Loc™ PBT Non-Absorbable Reload	VLOCN004L, VLOCN006L, VLOCN008L, VLOCN204L, VLOCN206L, VLOCN208L, VLOCN304L, VLOCN306L, VLOCN308L	Unexpired lots beginning with prefix "N0", "N1", "N9", "N2A", "N2B", "N2C", "N2D", and "N2E"

Customer Actions:

- Identify and quarantine all unused affected Endo Stitch Single Use Loading Unit with V-Loc Reloads.
- Return all unused affected product in your inventory to Medtronic as indicated in the Shipping and Return Instructions below.
- Please complete the Customer Acknowledgment Form even if you **do not** have unused inventory.
- Pass on this notice to all those who need to be aware within your organization or to any organization where the potentially affected devices have been transferred.

Shipping and Return Instructions:

	Customer with inventory	Customer with zero inventory	Where to send the completed form
Purchased directly from Medtronic	Please complete the attached Returns Verification Form in its entirety. Upon receiving your form, Medtronic Customer Care will contact you to organize the return of your products. You will receive credit for unused device(s) that you return	Complete form and check the box indicating "no inventory"	E-mail or fax the completed form to the Medtronic contact provided on the verification form.

Medtronic

	Customer with inventory	Customer with zero inventory	Where to send the completed form
Purchased from a distributor	Complete all fields on the form and contact your distributor directly to arrange for return of product.	Complete form and check the box indicating "no inventory"	E-mail or fax the completed form to your Distributor and to the Medtronic contact provided on the verification form.

Additional Information:

Medtronic has notified the Competent Authority of your country of this action.

We regret any inconvenience this may cause. We are committed to patient safety and appreciate your prompt attention to this matter. If you have any questions regarding this communication, please contact your Medtronic representative.

Sincerely,

Local / BU Manager

Enclosures:

Attachment A: IDENTIFYING AFFECTED PRODUCT

Attachment A:

IDENTIFYING AFFECTED PRODUCT

Endo Stitch™ Single use Loading Unit with V-Loc™ 180 Absorbable Reload

Endo Stitch™ Single Use Loading Unit with V-Loc™ PBT Non-Absorbable Reload

Product Name	Models	Lot Numbers
Endo Stitch V-Loc™ 180 Absorbable Reload	VLOCA004L, VLOCA006L, VLOCA008L, VLOCA204L, VLOCA206L, VLOCA208L, VLOCA304L, VLOCA306L, VLOCA308L	Unexpired all lots beginning with the "N0", "N1", "N9", "N2A", "N2B", "N2C", "N2D", and "N2E"
Endo Stitch V-Loc™ PBT Non-Absorbable Reload	VLOCN004L, VLOCN006L, VLOCN008L, VLOCN204L, VLOCN206L, VLOCN208L, VLOCN304L, VLOCN306L, VLOCN308L	Unexpired all lots beginning with the "N0", "N1", "N9", "N2A", "N2B", "N2C", "N2D", and "N2E"

COVIDIEN™

V-Loc™

PBT Non- Absorbable Reload

10 mm

0 3.5 Metric
4" 10 cm
BLUE / BLEU / BLAU / BLU / AZUL

ES-9

9 mm

REF VLOCN004L

LOT XXXXXXXXXXXX

Use by XXXX-XX

Model Number

Lot Number

Expiration Date

Polybutester (PBT) Nonabsorbable Knotless Wound Closure Device Reload For use with the Endo Stitch™ and SILS™ Stitch suturing devices
Recharge de dispositif de fermeture de plaie non résorbable sans nœud en polybutester (PBT)
Nachladevorrichtung für die nicht resorbierbare Polybutester-(PBT-)Wundverschlussvorrichtung ohne Knoten

Ricarica del dispositivo per chiusura di ferite con filo in polibutestere (PBT) non assorbibile senza nodi
Recarga no absorbible de polibutéster (PBT) para dispositivo de cierre de heridas sin nudos
Recarga do dispositivo de fechamento de ferida sem nó, não absorvível e composto de polibutéster (PBT)

© 2011 Covidien. Made In USA.
Covidien Inc, 15 Hampshire Street, Mansfield, MA 02048 USA.

STERILE EO

Single use

Caution, consult accompanying documents

CE 0123