



FSCA Ref: SAGQI-631, #1845810

FSN_with acknowledgment form_CS-104_SAGQI-631_EN.docx

Field Safety Notice (FSN)

CARDIOVIT CS-104

manufactured by

SCHILLER AG, Altgasse 68 CH-6341 Baar, Switzerland

www.schiller.ch

SRN: CH-MF-000012722 / CHRN: CHRN-MF-20000327

Date: 2023-02-28

Attention: SCHILLER AG authorized distributors and their customers

A problem related to the description of the electrical safety class in the Installation and Service manual of the device CARDIOVIT CS-104

This notice is intended to inform you about:

- what the problem is and under what circumstances it can occur.
- the actions that you as a distributor can take to minimize the effect of the problem.

We kindly ask that you read this notice carefully and send us written acknowledgement by **15th April 2023** that you have read and understood the contents of this notice. Written acknowledgement can be sent to SCHILLER AG to vigilance@schiller.ch.

SCHILLER AG apologizes for any inconveniences caused by this problem.

Sincerely,

Eckard Glaser

Head of Quality Management

Altgasse 68, CH-6341 Baar, Schweiz

vigilance@schiller.ch

Tel.: +41 41 766 42 42



FSCA Ref: SAGQI-631, #1845810
FSN_with acknowledgment form_CS-104_SAGQI-631_EN.docx

1. INFORMATION ON AFFECTED DEVICES	
COMMERCIAL NAME(S):	CARDIOVIT CS-104
PRIMARY CLINICAL PURPOSE OF DEVICE(S)*	The CARDIOVIT CS-104 is a 12-lead ECG and Spirometry (optional) system intended to be used by trained medical professionals in healthcare facilities for cardiological diagnosis in adult and paediatric patients. Analysis of the ECG and Spirometry is accomplished with algorithms that provide measurements, data presentations, graphical presentations, and interpretations for review by the user.
MODEL/CATALOGUE/ REF NUMBER(S):	0A.111000 (CARDIOVIT CS-104 (complete)) / 5.111001 (Standard SW CS-104) 0A.111010 (CARDIOVIT CS-104 (system complete)) / 5.111001 (Standard SW CS-104) 0A.111020 (CARDIOVIT CS-104 (UK version)) / 5.111001 (Standard SW CS-104) 0A.111030 (CARDIOVIT CS-104 (update set)) / 5.111001 (Standard SW CS-104) 0A.111040 (CARDIOVIT CS-104 (for replacement delivery)) / 5.111001 (Standard SW CS-104) 0A.111100 (CARDIOVIT CS-104 (Spiro complete)) / 5.111001 (Standard SW CS-104)
AFFECTED SERIAL OR LOT NUMBER RANGE:	All distributed devices.
UNIQUE DEVICE IDENTIFIER(S) (UDI-DI):	0A.111000 (CARDIOVIT CS-104 (complete)): - 5.111001 (Standard SW CS-104): 07613365003918 0A.111010 (CARDIOVIT CS-104 (system complete)): - 5.111001 (Standard SW CS-104): 07613365003918 0A.111020 (CARDIOVIT CS-104 (UK version)): - 5.111001 (Standard SW CS-104): 07613365003918 0A.111030 (CARDIOVIT CS-104 (update set)): - 5.111001 (Standard SW CS-104): 07613365003918 0A.111040 (CARDIOVIT CS-104 (for replacement delivery)): - 5.111001 (Standard SW CS-104): 07613365003918 0A.111100 (CARDIOVIT CS-104 (Spiro complete)): - 5.111001 (Standard SW CS-104): 07613365003918
DEVICE TYPE:	Interpretierender Mehrkanal-Elektrokardiograph
2. REASON FOR FIELD SAFETY CORRECTIVE ACTION (FSCA)	
PROBLEM DESCRIPTION	The description of the electrical safety test in the Installation and Service Instructions (Ref.: 2.540110, Rev: c) is described insufficiently and may lead to confusion regarding the electrical PROTECTION CLASS.
HAZARD GIVING RISE TO THE FSCA	The CARDIOVIT CS-104 (further called CS-104) is a software which is installed on a standalone PC/laptop, or on a PC incorporated in a device trolley. Installation of the CARDIOVIT CS-104 software is normally carried out by SCHILLER staff on site.



FSCA Ref: SAGQI-631, #1845810

FSN_with acknowledgment form_CS-104_SAGQI-631_EN.docx

	<p>For all configurations, an independent compatible SCHILLER recording device is used that can be positioned in different ways for patient convenience. Users (health care facilities) can install, assemble, combine, and configure the devices on their site according to their internal requirements. Depending on installation of the CS-104 and further assemblies, chosen by a health facility, this device can be used in combination with other medical and non-medical devices.</p> <p>An installation and assembling involving compatible SCHILLER devices is verified and tested for safety and compatibility by SCHILLER AG internally. The CARDIOVIT CS-104 can be used e.g., in the following combination: a PC, monitor, CARDIOVIT MS-12 blue or CARDIOVIT MS-12 USB or CARDIOVIT FT-1, and Trolley:</p> <ul style="list-style-type: none"> • SCHILLER medical devices / medical electrical equipment (e.g., CARDIOVIT MS-12 blue or CARDIOVIT MS-12 USB or CARDIOVIT FT-1), which can be used in combination with the installed CS-104 are classified as PROTECTION CLASS II or INTERNALLY POWERED. • A Trolley containing the PC and the power supplies is fully isolated. It is classified as a PROTECTION CLASS II. The internally used power supplies have an Earth connection, but this is NOT a protective earth it is a Functional Earth. All accessible metal parts during normal use are insulated. • A monitor (screen) is an ITE device and classified as a PROTECTION CLASS I equipment. The screen has accessible metal parts (connectors and screws). According to the IEC 60601-1 + AMD1:2012 + AMD2:2020, Chapter 16.1, the connections from the display and the display itself are “outside the patient environment”, and therefore, the requirements are sufficient according to the respective product IEC or ISO standards. <p>Listed above SCHILLER ME equipment is tested according to the applicable safety standard IEC 60601-1 and is safe for use. According to IEC 60601-1 + AMD1:2012 + AMD2:2020 clause 6 the ME EQUIPMENT can have multiple classifications. The screen is a Non-ME EQUIPMENT and shall comply with IEC safety standards that are relevant to that equipment. The used screen complies with: IEC 62368-1, IEC 62311, IEC 62479.</p> <p>That means the omission of the tests according to PROTECTION CLASS I for the combination are not leading to a hazardous situation nor a harm for the user or the patient. Therefore, a test according to PROTECTION CLASS I does not seem to be necessary to maintain patient and operator safety.</p>
<p>PREDICTED RISK TO PATIENT/USERS</p>	<p>Testing the device to PROTECTION CLASS II instead of PROTECTION CLASS I poses no risk to the patient or user.</p>



FSCA Ref: SAGQI-631, #1845810
FSN_with acknowledgment form_CS-104_SAGQI-631_EN.docx

3. TYPE OF ACTION TO MITIGATE THE RISK	
ACTIONS TO BE TAKEN BY THE USER	No actions to be taken by the USER.
ACTIONS TO BE TAKEN BY AUTHORIZED DISTRIBUTOR / IMPORTER	<ol style="list-style-type: none">1) Always use the current installation and service instructions and follow their instructions.2) If customers have been trained in service, provide them with the latest version of the installation and service handbook.3) Send the signed ANNEX I – Distributor/Importer Reply Form back to SCHILLER AG by 15th April 2023 as confirmation that the content of this notice was read and understood.
DATE FOR COMPLETION:	15th April 2023
IS THE FSN REQUIRED TO BE COMMUNICATED TO THE PATIENT / LAY USER?	No
FURTHER INFORMATION AND SUPPORT	If you need any further information, please send your request to vigilance@schiller.ch .

Transmission of this Field Safety Notice

This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)
Please transfer this notice to other organisations on which this action has an impact. (As appropriate)
Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.
Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback. *

The responsible National Authority has been informed about this communication of this field safety notice.

Contact person of manufacturer:

Eckard Glaser
Head of Quality Management
Altgasse 68, CH-6341 Baar, Switzerland
vigilance@schiller.ch
T: +41 41 766 42 42



FSCA Ref: SAGQI-631, #1845810
FSN_with acknowledgment form_CS-104_SAGQI-631_EN.docx

ANNEX I - Distributor/Importer Reply Form

1. Field Safety Notice (FSN) information	
FSN Reference number*	SAGQI-631
FSN Date*	2023-02-28
Product/ Device name*	CARDIOVIT CS-104

2. Manufacturer Details	
Company Name	SCHILLER AG
SRN	CH-MF-000012722
CHRN	CHRN-MF-20000327
Address	Altgasse 68 6341 Baar, Switzerland
Contact Name	Eckard Glaser
Email	vigilance@schiller.ch
Telephone Number	+41 41 766 42 42

3. Distributor/Importer Details	
Company Name*	
Account Number	
Address*	
Shipping address if different to above	
Contact Name*	
Title or Function	
Telephone number*	
Email*	

4. Distributors/Importers (Tick all that apply)		
<input type="checkbox"/>	*I confirm the receipt, the reading and understanding of the Field Safety Notice.	Distributor/Importer to complete or enter N/A
<input type="checkbox"/>	I have provided all service-trained customers with the latest version of the installation and service handbook.	Distributor/Importer to complete or enter N/A
Print Name*		
Signature*		
Date *		

Mandatory fields are marked with *

It is important that your organisation takes the actions detailed in the FSN and confirms that you have received the FSN.

Your organisation's reply is the evidence we need to monitor the progress of the corrective actions.