



Carl Zeiss Meditec AG 10589 Berlin

To whom it may concern

Division/Dept.: Complaint Management & Vigilance
Your contact: Dr. Lucia Puettmann, Claudia Minke

Carl Zeiss Meditec AG

Berlin location
Max-Dohrn-Strasse 8-10
10589 Berlin
Germany

Phone: N/A
Fax: N/A
E-mail: dl.med-complaints-lrb.all@zeiss.com

Your ref.: N/A
Yours of: N/A
Our ref.: FSCA BER 2025-01
Date: 2025-01-15

**URGENT/IMMEDIATE ACTION REQUIRED:
FIELD SAFETY CORRECTIVE ACTION (FSCA)
RECALL of Toric Intraocular Lenses (repacked units)**

Dear Customer,

You are using our intraocular lenses (IOL) and we thank you for your loyalty and trust in our products.

At ZEISS, the quality and safety of all our products is our highest priority. Unfortunately, with this letter, we must inform you, that we detected a possible labelling error on a production order of the above-mentioned IOLs and that we will therefore perform a Field Safety Corrective Action. In the following, we will give you a precise description of the situation and provide clear guidance on how to avoid any inconveniences for your patients.

Problem description:

During the repackaging process in production, it has been detected that some secondary labels showed an incorrect dataset for sphere and cylinder diopter. It is important to note that the primary label on the lens vial or barquette always contains the correct dataset of sphere and cylinder diopter.

We are performing repacking exclusively for toric IOLs; therefore, non-toric IOLs are not affected.

The potentially affected repacked toric IOLs have been identified. Until today we have not received complaints with any of these IOLs nor negative reports from dealers or SSCs. Nevertheless, it remains possible that a product with incorrect information on the secondary label has entered the market.

Address of Record:
Goeschwitzer Strasse 51 - 52
07745 Jena, Germany

Address for Delivery:
Carl Zeiss Meditec AG
Max-Dohrn-Strasse 8 - 10
10589 Berlin, Germany

Banks:
Deutsche Bank Jena
Account: 624536900 (BIC 820 700 00)
IBAN: DE90 8207 0000 0624 5369 00
BIC/ SWIFT: DEUT DE 8EXXX

Commerzbank Jena
Account: 258072800 (BIC 820 400 00)
IBAN: DE31 8204 0000 0258 0728 00
BIC/ SWIFT: COBADEFFXXX

Commercial Register:
Local Court Jena HRB 205623

VAT-ID No.: DE 811 922 737
WEEE-Reg.-No.: DE55298748

Chairman of the Supervisory Board:
Dr. Karl Lamprecht

Board of Management:
Dr. Markus Weber (CEO)
Justus Felix Wehmer

In consequence, we, Carl Zeiss Meditec AG, have decided to initiate a Field Safety Corrective Action for all IOLs identified, to inform customers and prevent further implantation of an IOL with the wrong diopter to avoid further harm to patients.

Affected products:

Our database indicates that you may have received one or multiple lenses referenced hereafter:

| Serial Number | IOL Description | Country |
|---------------|--|---------|
| 1S231344E019 | AT TORBI 709MP DPT -00.5 CYL 01.0 | AU |
| 1S232489E087 | AT LISA TRI TORIC 949MP SE+24.50 CYL01.5 | BE |
| 1S232505E281 | AT LISA TRI TORIC 949MP SE+23.50 CYL01.5 | BE |
| 1S232807E124 | AT LISA TRI TORIC 949MP SE+23.50 CYL01.0 | BE |
| 1S233080E021 | AT LISA TRI TORIC 949MP SE+23.50 CYL01.0 | BE |
| 1S232200E137 | AT TORBI 719MP SE+17.00 CYL02.5 | CH |
| 1S212124E038 | AT LISA TRI TORIC 949MP SE+21.00 CYL02.5 | DE |
| 1S212413E021 | AT TORBI 719M SE+22.50 CYL02.5 | DE |
| 1S212413E022 | AT TORBI 719M SE+22.50 CYL02.5 | DE |
| 1S223032E125 | AT LISA TRI TORIC 949MP SE+10.50 CYL03.5 | DE |
| 1S230725E135 | AT LISA TRI TORIC 949MP SE+23.50 CYL01.5 | DE |
| 1S230993E043 | AT LISA TRI TORIC 949MP SE+13.50 CYL02.5 | DE |
| 1S231094E236 | AT LISA TRI TORIC 949MP SE+18.00 CYL01.5 | DE |
| 1S231677E042 | AT LISA TRI TORIC 949MP SE+09.50 CYL02.5 | DE |
| 1S232031E053 | AT LISA TRI TORIC 949MP SE+26.00 CYL01.0 | DE |
| 1S232179E186 | AT TORBI 719MP SE+20.50 CYL01.5 | DE |
| 1S232461E059 | AT LISA TRI TORIC 949MP SE+19.00 CYL02.0 | DE |
| 1S232512E021 | AT LISA TRI TORIC 949MP SE+26.00 CYL01.5 | DE |
| 1S232600E038 | AT LISA TRI TORIC 949MP SE+25.00 CYL01.5 | DE |
| 1S232611E066 | AT LISA TRI TORIC 949MP SE+19.00 CYL01.0 | DE |
| 1S232626E032 | AT TORBI 719M SE+23.00 CYL08.0 | DE |
| 1S232672E051 | AT LISA TRI TORIC 949MP SE+09.00 CYL01.5 | DE |
| 1S232728E102 | AT LISA TRI TORIC 949MP SE+19.50 CYL01.5 | DE |
| 1S232872E029 | AT TORBI 719M SE+33.00 CYL03.5 | DE |
| 1S232880E066 | AT LISA TRI TORIC 949MP SE+26.00 CYL01.5 | DE |
| 1S232899E001 | AT LISA TRI TORIC 949MP SE+15.00 CYL01.5 | DE |
| 1S232899E008 | AT LISA TRI TORIC 949MP SE+15.00 CYL01.5 | DE |
| 1S232899E025 | AT LISA TRI TORIC 949MP SE+14.00 CYL02.0 | DE |
| 1S232899E069 | AT LISA TRI TORIC 949MP SE+13.50 CYL01.0 | DE |
| 1S232914E109 | AT LISA TRI TORIC 949MP SE+21.50 CYL01.5 | DE |
| 1S232955E122 | AT LISA TRI TORIC 949MP SE+16.50 CYL02.0 | DE |
| 1S233318E152 | AT LISA TRI TORIC 949MP SE+22.00 CYL01.5 | DE |
| 1S233491E059 | AT LISA TRI TORIC 949MP SE+25.00 CYL01.0 | DE |
| 1S233491E119 | AT LISA TRI TORIC 949MP SE+22.00 CYL01.5 | DE |

| Serial Number | IOL Description | Country |
|---------------|--|---------|
| 1S233615E044 | AT LISA TRI TORIC 949MP SE+15.50 CYL01.0 | DE |
| 1S233655E018 | AT TORBI 719M SE+25.50 CYL12.0 | DE |
| 1S233658E084 | AT LISA TRI TORIC 949MP SE+27.00 CYL01.0 | DE |
| 1S240199E171 | AT LISA TRI TORIC 949MP SE+25.00 CYL01.0 | DE |
| 1S240226E094 | AT LISA TRI TORIC 949MP SE+20.50 CYL01.0 | DE |
| 1S240341E076 | AT LISA TRI TORIC 949MP SE+22.00 CYL01.5 | DE |
| 1S240412E082 | AT LISA TRI TORIC 949MP SE+21.50 CYL01.0 | DE |
| 1S240496E034 | AT LISA TRI TORIC 949MP SE+24.00 CYL01.0 | DE |
| 1S240711E022 | AT TORBI 719MP SE+02.50 CYL02.0 | DE |
| 1S240894E133 | AT LISA TRI TORIC 949MP SE+27.00 CYL01.0 | DE |
| 1S240188E083 | AT LISA TRI TORIC 949MP SE+20.00 CYL01.0 | DK |
| 1S240381E130 | AT LISA TRI TORIC 949MP SE+20.00 CYL01.0 | DK |
| 1S240894E103 | AT LISA TRI TORIC 949MP SE+20.00 CYL01.5 | DK |
| 1S212388E080 | AT TORBI 719M SE+15.50 CYL02.5 | ES |
| 1S212614E085 | AT TORBI 719M SE+25.50 CYL05.5 | ES |
| 1S231951E203 | AT LISA TRI TORIC 949MP SE+21.50 CYL02.5 | ES |
| 1S232098E212 | AT TORBI 719MP SE+15.50 CYL03.5 | ES |
| 1S240118E049 | AT LISA TRI TORIC 949MP SE+19.50 CYL02.0 | ES |
| 1S241425E016 | AT TORBI 719MP SE+05.00 CYL02.5 | ES |
| 1S212614E002 | AT TORBI 719M SE+21.50 CYL05.0 | FR |
| 1S212638E025 | AT TORBI 719M SE+25.50 CYL02.5 | FR |
| 1S231363E035 | AT TORBI 719MP SE+18.00 CYL02.5 | FR |
| 1S231963E009 | AT TORBI 719MP SE+18.00 CYL02.5 | FR |
| 1S232383E216 | AT TORBI 719MP SE+28.00 CYL02.0 | FR |
| 1S232872E102 | AT TORBI 719M SE+17.50 CYL04.5 | FR |
| 1S233261E026 | AT TORBI 719MP SE+17.00 CYL02.5 | FR |
| 1S233310E012 | AT TORBI 719M SE+18.50 CYL06.5 | FR |
| 1S233560E078 | AT TORBI 719M SE+29.00 CYL04.5 | FR |
| 1S241026E027 | AT TORBI 719M SE+16.00 CYL06.0 | FR |
| 1S241825E016 | AT TORBI 719M SE+22.00 CYL08.0 | FR |
| 1S231589E156 | AT TORBI 709MP DPT 20.5 CYL 02.0 | MY |
| 1S233096E005 | AT LISA TRI TORIC 949MP SE+19.50 CYL01.5 | NO |
| 1S231241E090 | AT LISA TRI TORIC 949MP SE+25.00 CYL01.5 | PT |
| 1S212404E004 | AT LISA TRI TORIC 949MP SE+21.00 CYL02.5 | RO |
| 1S230915E030 | AT LISA TRI TORIC 949MP SE+24.00 CYL03.5 | RO |
| 1S232430E166 | AT LISA TRI TORIC 949MP SE+11.50 CYL02.5 | RO |
| 1S232430E170 | AT LISA TRI TORIC 949MP SE+11.50 CYL02.5 | RO |
| 1S232512E096 | AT LISA TRI TORIC 949MP SE+30.00 CYL02.5 | RO |

| Serial Number | IOL Description | Country |
|---------------|--|---------|
| 1S233615E057 | AT LISA TRI TORIC 949MP SE+26.00 CYL01.0 | RO |
| 1S231431E186 | AT TORBI 709MP DPT -03.0 CYL 01.0 | ID |
| 1S221506E314 | AT TORBI 709MP DPT 18.0 CYL 02.0 | TH |

Hazard description

Inaccurate information on the secondary label may lead to implantation of a lens with an incorrect power and the patient may have an unintended outcome of refraction as well as visual acuity.

In case of a wrong refraction results, an additional surgery may be required to correct the error, based on your judgement of the benefit / risk for the patient:

- either an explantation/reimplantation of a new IOL,
- or a secondary IOL implantation in sulcus,
- or an additional refractive surgery,
- or eyeglasses/contact lenses correction prescription.

Actions & Recommendation:

Please check the status of all affected products you have:

- If you have still one of these lenses in stock, please place them immediately in quarantine and contact your local ZEISS representative. These lenses must be shipped back to ZEISS.
- If you have already implanted one of the listed devices, please review the diopter information on the label in the patient's file which is giving the correct information about the IOL's power. In case the diopter on this label is not the intended IOL power, please review the refractive outcome of the patient.

Please inform the relevant persons within your healthcare structure who are involved in use of the above-mentioned ZEISS intraocular lenses.

We kindly ask you to send back to us the acknowledge receipt of the letter which you will find in Appendix 1.

This Field Safety Corrective Action will be reported to your local Health Authorities in accordance with local regulations.

We thank you for your careful attention, your consequent verifications, and your continuous support. We sincerely regret the inconvenience caused and thank you for addressing the matter promptly. We remain at your disposal.

Yours sincerely,

Carl Zeiss Meditec AG
i.V.

Carl Zeiss Meditec SAS
i.V.

Dr. Lucia Puettmann
Head of Complaint Management & Vigilance
ZEISS Medical Technology Segment

Claudia Minke
Complaint & Vigilance Manager
Implants & Disposables
ZEISS Medical Technology Segment

Annex

Appendix 1: Confirmation sheet

RECALL FSCA BER 2025-01

I have read and understood the RECALL information related to FSCA BER 2025-01.

I have transmitted the information to the relevant persons within my healthcare structure.

Status of the affected lenses:

| Product Name and Diopter (D) | Serial Number(s) | Lens Status: <ul style="list-style-type: none"> • Blocked/Sent back to ZEISS • Implanted/Patient outcome |
|-------------------------------------|-------------------------|--|
| | | |

Confirmation:

Signature: _____ Date: _____

| | |
|-----------------|--|
| Name: | |
| Function: | |
| Address: | |
| Phone: | |
| E-mail address: | |

Please send back this confirmation form via e-mail to

- dl.med-complaints-lrb.all@zeiss.com