

Rev 2: February 2020

FSN Ref: CAPA-2025-05

FSCA Ref: CAPA-2025-05

Date: 2025-03-04

Field Safety Notice
MICROCONE NI implant

For Attention of*: FSCA responsible person in your organization

Contact details of local representative (name, e-mail, telephone, address etc.)*

This could be a distributor or local branch of the manufacturer. To be added at the appropriate stage in the different local languages.

Field Safety Notice (FSN)
MICROCONE NI implant
Risk addressed by FSN

1. Information on Affected Devices*			
1.	1. Device Type(s)*		
	MICROCONE implants are screw-shaped dental implants which serve as tooth root replacements.		
1.	2. Commercial name(s)*		
	MICROCONE implant		
1.	3. Unique Device Identifier(s) (UDI-DI)		
	04251574831859 04251574831866 04251574831873		
1.	4. Primary clinical purpose of device(s)*		
	Restoration of the esthetics and functionality of the teeth in partially or fully edentulous patients. The MICROCONE NI Implants D 3.0 may only be used to replace the upper, lateral incisors (region 12/22) and the lower, middle and lateral incisors (region 31/32/41/42).		
1.	5. Device Model/Catalogue/part number(s)*		
	1-01-06 1-01-07 1-01-08		
1.	6. Software version		
	n/a		
1.	7. Affected serial or lot number range		
	1-01-06	1-01-07	1-01-08
	L0063582	L0063593	L0063596
	L0065343	L0064678	L0065345
	L0072370	L0065406	L0076434
	L0073077	L0065408	L0090086
	L0073078	L0075120	L0090087
	L0074465	L0076079	L0116201
	L0075119	L0077622	/
	L0075357	L0085095	/
	L0076433	L0085096	/
	L0077621	L0090005	/
	L0078616	L0090998	/
	L0082746	L0116199	/
	L0082747	L0116200	/
	L0083949	L0117450	/
	L0083951	L0124710	/
	L0090543	/	/
	L0093926	/	/
	L0098542	/	/
	L0103386	/	/

	L0110966			
	L0110968			
	L0114579			
	L0116197			
	L0119513			
	L0123340			
	L0123341			
1.	8. Associated devices			
	n/a			

2. Reason for Field Safety Corrective Action (FSCA)*	
2.	1. Description of the product problem* It has been detected that the above listed batches are manufactured from titanium grade 4 instead of Ti-6Al-4V ELI.
2.	2. Hazard giving rise to the FSCA* MICROCONE NI implants with diameter D 3.0 are used to replace the upper, lateral incisors (region 12/22) and the lower, middle and lateral incisors (region 31/32/41/42). In comparison to MICROCONE RI implants with diameter D 3.5-5.0 which are manufactured from titanium grade 4, MICROCONE NI implants with diameter D 3.0 are manufactured from the material Ti-6Al-4V ELI, whereas both materials are SOTA dental implant materials. Complaint data available to date do not indicate a significant difference between the conforming and nonconforming products. Our investigation including biocompatibility assessment or stability testing shows that there is no elevated risk for the patient in comparison to the conforming product. However, to prevent further implantation of nonconforming products, the affected batches are recalled. Already successfully implanted nonconforming products should be monitored at regular intervals in the dental practice and do not need to be removed or replaced.
2.	3. Probability of problem arising -
2.	4. Predicted risk to patient/users -
2.	5. Further information to help characterise the problem -
2.	6. Background on Issue -
2.	7. Other information relevant to FSCA -

3. Type of Action to mitigate the risk*	
3.	1. Action To Be Taken by the User* <input checked="" type="checkbox"/> Identify Device <input checked="" type="checkbox"/> Quarantine Device <input checked="" type="checkbox"/> Return Device
3.	2. By when should the action be completed? <div style="text-align: center; background-color: yellow;">24 MAR 2025</div>

3.	3. Particular considerations for: Implantable device	
	Is follow-up of patients or review of patients' previous results recommended? No Already successfully implanted nonconforming products should be monitored at regular intervals in the dental practice and do not need to be removed or replaced.	
3.	4. Is customer Reply Required? * (If yes, form attached specifying deadline for return)	Yes
3.	5. Action Being Taken by the Manufacturer*	
	<input checked="" type="checkbox"/> Product Removal Affected products on the market will be recalled.	
3.	6. By when should the action be completed?	04 SEP 2025
3.	7. Is the FSN required to be communicated to the patient /lay user?	No
3.	8. If yes, has manufacturer provided additional information suitable for the patient/lay user in a patient/lay or non-professional user information letter/sheet?	
	Choose an item.	Choose an item.

4. General Information*	
4.	1. FSN Type* New
4.	2. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)
	a. Company Name MEDENTIKA GmbH
	b. Address Hammweg 8-10, 76549 Hügelsheim, Germany
	c. Website address www.medentika.de
4.	3. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. *
4.	4. List of attachments/appendices: If extensive consider providing web-link instead.
4.	5. Name/Signature Insert Name and Title here and signature below.

Transmission of this Field Safety Notice	
	<p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate)</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate)</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.*</p>

Note: Fields indicated by * are considered necessary for all FSNs. Others are optional.