

Urgent Field Safety Notice

ACHC26-03.A.OUS

Atellica CH Analyzer

Atellica CI Analyzer

Title	Atellica CH Urine Albumin (UA1b) – Potential for Falsely Depressed Results		
Date Issued	JAN-2026		

Product	Assay	Test Code	Siemens Material Number / Unique Device Identification	Lot Number
	Atellica CH Urine Albumin (UA1b)	UA1b	11537225/ 00630414611099	All lots

Issue Description	Siemens Healthineers has confirmed, through investigation of customer complaints, a potential for falsely depressed results for a small subset of samples above the measuring interval when using the Atellica CH Urine Albumin (UA1b) assay on the Atellica CH and Atellica CI analyzers. The investigation determined that certain samples may report results within the UA1b measuring interval when expected to be above the UA1b measuring interval, when compared to the Atellica CH Microalbumin_2 (μ ALB_2) assay result. Based on global customer data and internal analysis, this subset represents approximately 0.3% of samples. All current and future UA1b lots are considered impacted until further notice. Siemens Healthineers is actively investigating the root cause of this issue.
--------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Impact to Results	Falsey depressed UA1b patient results may occur and may be reported due to this issue. In this case, affected samples with any urine albumin concentration above the UA1b measuring interval ($>40.0 \text{ mg/dL}$; $>400 \text{ mg/L}$) may report values between 6.2 mg/dL (62 mg/L) and 40.0 mg/dL ($>400 \text{ mg/L}$). Results of this assay should always be interpreted in conjunction with the patient's medical history, clinical presentation, and other findings.
--------------------------	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Customer Actions	<ul style="list-style-type: none">Internal investigation has confirmed that diluting mitigates this behavior. A x10 dilution of an affected sample produces accurate UA1b results.As a temporary mitigation, for any sample with a UA1b result above 5.0 mg/dL (50 mg/L), reanalyze the sample with a x10 dilution.<ul style="list-style-type: none">If the diluted result is consistent with the initial result, either result may be reported.If the diluted result is inconsistent with the initial result, the diluted result should be reported.If you prefer to automate this dilution, please refer to your internal laboratory procedures.Please review this notification with your Medical Director to determine the appropriate course of action, including evaluation of any previously generated results, if applicable.Retain this letter with your laboratory records and forward it to others who may have received this product.
-------------------------	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Resolution	A follow-up communication will be provided when “Customer Actions” are no longer required.
Single Registration Number (SRN)	US-MF-000016560
<hr/> <p>We appreciate your continued partnership and apologize for the inconvenience this situation may cause. If you have any questions or require assistance, please contact your Siemens Healthineers Customer Care Center or your local Siemens Healthineers technical support representative.</p> <hr/>	

All trademarks are the property of their respective owners